

Kansas Administrative Regulations

Kansas Board of Examiners in Fitting and Dispensing of Hearing Instruments

Article 7.—Calibration of Audiometric Equipment

67-7-4. Calibration of audiometric equipment.

(a) Any audiometric equipment that is used in connection with the sale or fitting of hearing aids in this state shall be calibrated at intervals not exceeding two years.

(b) Each licensee shall submit to the board written proof of each calibration required in subsection (a) of this rule and regulation within 30 days of the required dates of calibration. Written proof of calibration shall include the following:

- (1) The name of the owner;
- (2) the make, model, and serial number of the equipment;
- (3) the date of calibration;
- (4) the printed name of the person and company calibrating the equipment;
- (5) the signature of the person calibrating the equipment; and
- (6) the name of the licensee submitting proof of calibration.

(c) Any of the following acts may constitute unethical conduct for which a licensee may be disciplined under K.S.A. 74-5818 and amendments thereto:

- (1) Provides false or misleading information under this rule and regulation;
- (2) uses audiometric equipment that has not been calibrated as required by this rule and regulation; or
- (3) uses audiometric equipment the calibration of which has not been reported as required by this rule and regulation.

(Authorized by and implementing K.S.A. 74-5806 and K.S.A. 1999 Supp. 74-5818; effective, T-86-16, June 17, 1985; effective May 1, 1986; amended May 12, 2000.)

***** Authenticated Kansas Administrative Regulation *****