

Kansas Administrative Regulations

Department of Health and Environment

Article 35.—Radiation

28-35-230f. Reports of leaking or contaminated sealed sources.

If a test for leakage or contamination pursuant to K.A.R. 28-35-216a indicates a sealed source is leaking or contaminated, a report of the test shall be filed within five days with the department describing the equipment involved, the test results and the corrective action taken.

(Authorized by and implementing K.S.A. 1993 Supp. 48-1607; effective Oct. 17, 1994.)

***** Authenticated Kansas Administrative Regulation *****