Kansas Administrative Regulations

Department of Health and Environment

Article 35.—Radiation—Part 9. Radiation Safety Requirements for Particle Accelerators.

28-35-311. Human use of particle accelerators.

In addition to the requirements set forth in part 2, the registrant shall: (A) Whenever deemed necessary by the department, the registrant shall appoint a medical committee of at least three (3) members to evaluate all proposals for research, diagnostic, and therapeutic use of a particle accelerator. Membership of the committee should include physicians expert in internal medicine, hematology, therapeutic radiology, and a person experienced in depth dose calculations and protection against radiation;

(B) The individuals designated as the users shall have substantial training and experience in deep therapy techniques or in the use of particle accelerations to treat humans; and

(C) The user must be a physician.

(Authorized by K.S.A. 1975 Supp. 48-1607; effective May 1, 1976.)

***** Authenticated Kansas Administrative Regulation *****